

Mise en service le 16/02/2024

Serial number: 620 001 577

Class of accuracy: 1

Measuring range: 0 - 800 bar

Test date: 13.02.2023

Name of tester: Mathias Klaus

Used test medium: Keller Manometer Typ LEO 1 / 1000bar / 81000.5

Serial number: 41551

Class of accuracy: 0,1

Area of display: 0.0 ... 1000.0bar

Certificate no.: 620 001 577 13-02-23

Gauge value	Measured value	Divergence
0 bar	0 bar	0 bar
100 bar	98 bar	-2 bar
200 bar	200 bar	0 bar
300 bar	303 bar	3 bar
400 bar	402 bar	2 bar
500 bar	496 bar	-4 bar
600 bar	601 bar	1 bar
700 bar	697 bar	-3 bar

valid range ± 8 bar

-8 bar - 8 bar

92 bar - 108 bar

192 bar - 208 bar

292 bar - 308 bar

392 bar - 408 bar

492 bar - 508 bar

592 bar - 608 bar

692 bar - 708 bar