

Jet Ro Air 30000 29416

HYTORC®

Technologies

Certificate of calibration

This calibration certificate documents the traceability to national standards which realize the physical units of measurement according to the international system of units (SI); traceability maintained through calibration report 81100076-20240603. Pressure gauge was calibrated on Gauge Calibration Case MPK0.1, serial number 81100076. Calibration performed at 20°C/68°F and 65% relative humidity (calibrated on 06/03/2024, calibration due date 09/02/2024).

Gauge type: Gauge RKG 100-1r 0-800bar/0-11500psi 1/4"NPT - calibrated
Serial number: HTG081568
Range: 0 - 11500 PSI / 0 - 800 BAR
Accuracy: 1% FS
Maximum permissible deviation: +/- 115,00 PSI

Gauge reading [PSI] / [BAR]	Applied pressure [PSI] / [BAR]	Deviation [PSI] / [BAR]
1000 / 68,9	987 / 68,1	-13 / -0,8
2000 / 137,9	1956 / 134,9	-44 / -3,0
3000 / 206,8	2993 / 206,4	-7 / -0,4
4000 / 275,8	3976 / 274,1	-24 / -1,7
5000 / 344,7	4970 / 342,7	-30 / -2,0
6000 / 413,7	5958 / 410,8	-42 / -2,9
7000 / 482,6	6953 / 479,4	-47 / -3,2
8000 / 551,6	7966 / 549,2	-34 / -2,4
9000 / 620,5	8965 / 618,1	-35 / -2,4
10000 / 689,5	9968 / 687,3	-32 / -2,2

Calibration status: **PASS**

Calibrations are performed in reference to ISO 10012:2003. Measurements performed in psi. Results are valid only for the above item calibrated on time of test. This certificate shall not be reproduced except in full without written permission of HYTORC.



Dino Lindrath (Tester)


M. Westermann (Quality assurance)

08/14/2024
Date of calibration (MM/DD/YYYY)

17/03/2025
Date pressure gauge put in service (MM/DD/YYYY)

Issued by: HYTORC Technologies GmbH, Klenzendorf 3-17, 45519 Sprockhövel, Tel. +49 2324 9793-0, Fax +49 2324 84753 095, info@hytorc.com


Guy Goux

HYTORC-HuSiach
ZA Chemin de Montépy
FLEURY-LES-BOIS
69210
Tél. : 04 78 33 39 19
E-mail : sn@hytorc-ce.com
N° SIRET : 808 881 189 00019
TVA : FR 11 808 881 189